



The temperature of the road surface is of great importance to the road authorities and road maintenance people, especially during the winter when snow and frost make it risky to travel along our highways. Having access to the road temperature at all times, those responsible can decide what action to take to keep the roads free from ice and snow.

The Road Temperature Sensor 3304 is designed to be embedded in the road tarmac with a fixed durable 15 meter cable and extension cable 2842 that connects the sensor to the Datalogger 3660 in the Road Weather Station nearby. The sensing element is a platinum resistor, Pt 2000 with positive temperature coefficient. The Pt element, together with 2 resistors, forms a Wheatstone half-bridge which is molded into a mixture of polyurethane and cement .

As the sensor is flush with the road surface it must wear at the same rate as the road tarmac. A vertical polystyrene rod is molded into the sensor for wear indication. When the rod appears on the surface of the sensor, after approximately 12mm wear, the sensor must be replaced.

Specifications

ROAD TEMPERATURE SENSOR 3304



Illustration shows the sensor embedded in the road

CIRCUIT DIAGRAM

Plug, pin =•; bushing =0



Sensing element: Range: Resolution: Accuracy:	Platinum Pt 2000 - 43 to + 48°C 0.1°C ± 0.2°C
Output:	VR22 (analog)
Time Constant (63%):	30 seconds
Current Consumption:	70µA during measurement
Electrical Connection:	15m fixed cable with 6-pin plug (Other lengths on request)
Accessories (optional):	Connecting Cable 2842
Wear Indicator:	Replace sensor when wear indicator is visible (Approx. 12mm wear). See illustration
Material:	Foss-Than 646 / Cement MP-30
Weight:	1100 grams with 15 meter cable

CALIBRATION

Temperature (°C)	Raw data N

To convert raw data to engineering units use the formula:

Temperature (°C) = $A + BN + CN^2 + DN^3$

Serial No: Calculated coefficients:

А	С	
В	D	0

Date:

Sign:



Visit our Web site for the latest version of this document and more information **www.aadi.no**

Aanderaa is a trademark of Xylem Inc. or one of its subsidiaries. © 2012 Xylem, Inc. D247 January 2013 Aanderaa Data Instruments AS Sanddalsringen 5b, P.O. Box 103 Midtun, 5828 Bergen, Norway Tel +47 55 60 48 00 Fax +47 55 60 48 01